

Analysis - SOLID	Preservation	Glass(G)/Plastic(P)	Size	Holding Time	Combinations/Notes
ABN/PAH 8270/625	None	G-Amber	4oz	14 days	These analyses may be taken from a 4oz amber jar. Analyses which can go into clear jars, may also share a 4oz Amber jar with the parameters to the left. * UP TO 4 analyses per jar. TCLP may require an extra jar.
TPH 8100/DRO 8015	None	G-Amber	4oz	14 days	
Pesticides 8081	None	G-Amber	4oz	14 days	
EPH	None	G-Amber	4oz	14 days	
Ignitability	None	G-Clear	4oz	14 days	Must have it's own container
PCB 8082	None	G-Clear	4oz	1 year	Up to 4 analyses can share one 4oz clear jar  If TCLP is required an additional 4oz jar is needed  Note: Analyses which require amber jars cannot be taken from clear jars
Anions( NO <sub>2</sub> , NO <sub>3</sub> ,Cl, Br, O-PO <sub>4</sub> , SO <sub>4</sub> )	None	G-Clear	4oz	NO <sub>2</sub> ,NO <sub>3</sub> ,O-PO <sub>4</sub> 7days All other anions 28days	
Hexavalent Chromium	None	G-Clear	4oz	30 days	
COD	None	G-Clear	4oz	28 days	
Conductivity	None	G-Clear	4oz	28 days	
Cyanide	None	G-Clear	4oz	14 days	
Metals	None	G-Clear	4oz	180 days, Hg 28 days	
pH	None	G-Clear	4oz	7 days	
Sulfide	None	G-Clear	4oz	7 days	
TS	None	G-Clear	4oz	7 days	
TKN	None	G-Clear	4oz	28 days	
Ammonia (NH <sub>3</sub> )	None	G-Clear	4oz	28 days	
Total Phosphorus	None	G-Clear	4oz	28 days	
VOC VPH	MeOH	G-Clear	40mL	28 days	
VOC 8260/8021/8015GRO	MeOH	G-Clear	40mL	14 days	

Proprietary Use of Absolute Resource Associates  
124 Heritage Avenue #16 | Portsmouth, NH 03801 | 603.436.2001  
absoluteresourceassociates.com

QSD-7 rev7 01/21/21 JVG